

CERTIFICATE OF ANALYSIS

ANKOM TDF Check Sample

Lot#: C-2023-001

Expiration Date: December 31, 2025

Cereal Check Sample (50g)

Total (%)

IDF: 14.03%, SD: 0.48

SDF: 2.09%, SD: 0.31

TDF: 16.15%, SD: 0.62

Residue (g)

IDF: 0.0841g, SD: 0.0015

SDF: 0.0247g, SD: 0.0026

TDF: 0.1088g, SD: 0.0025

Protein (g)

IDF: 0.0148g, SD: 0.0016

SDF: 0.0091g, SD: 0.0014

TDF: 0.0232g, SD: 0.0024

Ash (g)

IDF: 0.0060g, SD: 0.0006

SDF: 0.0082g, SD: 0.0009

TDF: 0.0134g, SD: 0.0017

The above sample was tested **“as received”** for the above analyses in the laboratory at ANKOM Technology using ANKOM Dietary Fiber (TDF).

This check sample is to only be used as an instrument troubleshooting tool. This is not a certified reference material. ANKOM considers data falling within 3 SD to indicate an instrument is performing correctly.

SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range. 95% of data points will be within 2 SD, 99% of data points will be within 3 SD.