

## CERTIFICATE OF ANALYSIS

### ANKOM Fat Check Sample

**Lot#: OM-2023-001**

**Check Sample (50g)**

Expiration Date: December 31, 2025

**Crude Fat:** 2.33%, SD: 0.14

**Moisture:** 6.06%, SD: 0.08

**Total Fat:** 6.45%, SD: 0.30

This check sample is to only be used as an instrument troubleshooting tool. This is not a certified reference material. ANKOM considers data falling within 3 SD to indicate an instrument is performing correctly.

SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range. 95% of data points will be within 2 SD, 99% of data points will be within 3 SD.