

CERTIFICATE OF ANALYSIS

ANKOM Fat Check Sample

Lot#: OM-2023-002

Check Sample (50g)

Expiration Date: December 31, 2026

Crude Fat: 3.61%, SD: 0.45

Moisture: 5.97%, SD: 0.30

Total Fat: 6.56%, SD: 0.75

This check sample is to only be used as an instrument troubleshooting tool. This is not a certified reference material. ANKOM considers data falling within 3 SD to indicate an instrument is performing correctly.

SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range. 95% of data points will be within 2 SD, 99% of data points will be within 3 SD.