


## Controller (Temp) Menu Set-up DELTA/DELTAI and A2000/A2000I

Service Procedure 015

Revised: 05/06/2021e

Your DELTA and A2000 will have one of the controllers shown here (ASCON and ATHENA). To configure, follow the steps below.



1. Power on the instrument.

## For ASCON:

After you power on the controller, if "ConF" is displayed, execute steps 4-26. If ambient temperature is displayed, execute steps 8-26. If you make a mistake at any point in the procedure, go back to step 8 and start over.

## For ATHENA:

After you power on the controller, you will need to enter a passcode. Therefore, you must execute steps 2-26. If you make a mistake at any point in the procedure, go back to step 2 and start over.

