

CERTIFICATE OF ANALYSIS

ANKOM TDF Check Sample

Lot#: C-2022-001

Expiration Date: December 31, 2024

Cereal Check Sample (50g)

IDF: 33.98%, SD: 0.43

SDF: 2.35%, SD: 0.62

TDF: 36.36%, SD: 0.51

Residue:

IDF: 0.1825g, SD: 0.0022

SDF: 0.0266g, SD: 0.0015

TDF: 0.2092g, SD: 0.0023

Residue Ash:

IDF: 0.0072g, SD: 0.0011

SDF: 0.0071g, SD: 0.0026

TDF: 0.0118g, SD: 0.0035

Residue Protein:

IDF: 0.123g, SD: 0.0008

SDF: 0.0107g, SD: 0.0011

TDF: 0.0225g, SD: 0.0008

The above sample was tested **“as received”** for the above analyses in the laboratory at ANKOM Technology using ANKOM Dietary Fiber (TDF).

This check sample is to only be used as an instrument troubleshooting tool. This is not a certified reference material. ANKOM considers data falling within 3 SD to indicate an instrument is performing correctly.

SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range. 95% of data points will be within 2 SD, 99% of data points will be within 3 SD.