

CERTIFICATE OF ANALYSIS

ANKOM TDF Check Sample

Lot#: C-2022-002

Expiration Date: December 31, 2024

Cereal Check Sample (50g)

Total (%)

IDF: 12.78%, SD: 0.59

SDF: 3.20%, SD: 0.48

TDF: 15.54%, SD: 0.51

Residue (g)

IDF: 0.0694-0.0829

SDF: 0.0167-0.0400

TDF: 0.0648-0.1355

Protein (g)

IDF: 0.0090-0.0190

SDF: 0.0000-0.0214

TDF: 0.0165-0.0319

Ash (g)

IDF: 0.0016-0.0089

SDF: 0.0037-0.0077

TDF: 0.0010-0.0151

The above sample was tested “**as received**” for the above analyses in the laboratory at ANKOM Technology using ANKOM Dietary Fiber (TDF).

This check sample is to only be used as an instrument troubleshooting tool. This is not a certified reference material. ANKOM considers data falling within 3 SD to indicate an instrument is performing correctly.

SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range. 95% of data points will be within 2 SD, 99% of data points will be within 3 SD.