

1. Installed part # 5806 O-Ring is in the top of the Vessel Insert.



2. The large O-Ring is installed in the Vessel as shown.



Note for large O-Ring:

Part #969.5 for Standard Pet  
Ether O-Ring

Part # 5663 for DiEthyl Ether  
(High Penetrating Solvents )