



Ankom Technology
2052 O'Neil Road
Macedon, NY 14502
(315) 986-8090

CERTIFICATE OF ANALYSIS

Chemical Name: Ankom Crude Fiber Base 0.313N

Ankom Part #: FCFB20CB

CAS No.: Mixture

LOT: AD-26181-8

Expiration Date: June 2029

	<u>SPECIFICATION</u>	<u>RESULT</u>
Appearance	Clear Liquid	Conforms
Titration using 0.1M HCl	0.308 – 0.318 N	0.315 N

Approved By: _____

A handwritten signature in black ink, appearing to be "S. Jackson", written over a horizontal line.

Samantha Jackson
Laboratory Manager

Date: _____

6/30/2020

