

Certificate of Analysis

Diatomaceous Earth Part #'s: DE1, DE2

Lot Number: 3C02B21D

Expiration:

31-Dec-24

<u>Test</u> Moisture Ash Test Result less than 0.15% * not less than 99.7% **

* Moisture was determined by drying at 100°-105° C for 3 hours

** Ash was determined by by placing in furnace at 600° C for 2 hours

Product Information

Contents: Amorphous diatomaceous earth and crystaline silica in various amounts Storage: Store at room temperature in closed container Country of Origin: United States of America

2052 O'Neil Rd., Macedon, NY 14502

www.ankom.com