

CERTIFICATE OF ANALYSIS

ANKOM Fiber Check Sample

Lot#: ALF-2023-001

Alfalfa Check Sample (50g)

Expiration Date: December 31, 2026

Crude Fiber: 25.9%, SD: 1.21

ADF: 28.10%, SD: 0.945

NDF: 32.32%, SD 0.477

*SD: Standard Deviation provided represents 1 SD from the mean, 68% of data points will statistically fall within this range.

*95% of data points will within 2 SD

The above sample was tested “**as received**” for the above analyses in the laboratory at ANKOM Technology using ANKOM Delta.