

CERTIFICATE OF ANALYSIS

ANKOM High Fat Check Sample

Lot#: 2022

Crisco Check Sample (50g)

Expiration Date: December 31, 2024

Fat: 99.5 % \pm 0.20

Moisture: 0.10 % \pm 0.05

Sample dried for 3 hours at 100° - 105°C

Sample extracted with petroleum ether

Sample is guaranteed for one year after receiving when stored at room temperature