



CERTIFICATE OF ANALYSIS

ANKOM A100 Alfalfa Check Sample January 1, 2022

ADF	33.5 % ± 0.8
NDF.....	38.5 % ± 0.8
Crude Fiber	28.4 % ± 0.7

The above sample was tested for the above analyses in the laboratory at ANKOM Technology using ANKOM fiber analyzers.