

Controller (Temp) Menu Set-up DELTA/DELTAI

Service Procedure 015

Revised: 04/26/22 RJCe

Your DELTA will have an ASCON controller shown here. To configure, follow the steps below.



1. Power on the instrument.

For ASCON:

After you power on the controller, if "ConF" is displayed, execute steps 4-26. If ambient temperature is displayed, execute steps 8-26. If you make a mistake at any point in the procedure, go back to step 8 and start over.

2.	Press	tab until you see	PASS on the display.			
3.	Press	until you see	33.0 on the display.	Press — You w	rill see ConF on the display.	
4.	Press	until you see	0101 on the display.	Press You w	vill see Un,t on the display.	
5.	Press	until you see	°C on the display.	Press You w	retr on the display.	
6.	Press	until you see	4-20 on the display.	Press You w	vill see O.c.rb on the display.	
7.	Press	until you see	0.5 on the display.	Press You w	vill see Code on the display.	
8.	Press	tab until you see	Pb. on the display.			
9.	Press	until you see	1.0 on the display.	Press — You w	vill see t.,. on the display.	
10.	Press	until you see	3>.0 on the display.	Press You w	vill see t.d. on the display.	
11.	Press	until you see	0.80 on the display.	Press — You w	vill see t.c. on the display.	
12.	Press	until you see	10 on the display.	Press — You w	vill see O.C. on the display.	
13.	Press	until you see	0.50 on the display.	Press You w	vill see OP. H on the display.	
14.	Press	until you see	100.0 on the display.			
15.	Press	tab until you see	tune on the display.			
16.	Press	until you see	StoP on the display.	Press — You w	rill see SL. U on the display.	
17.	Press	until you see	OFF on the display.	Press — You w	rill see SL. D on the display.	
18.	Press	until you see	OFF on the display.	Press — You w	rill see S.P. L on the display.	
19.	Press	until you see	0.0 on the display.	Press — You w	rill see S.P. H on the display.	
20.	Press	until you see	100.0 on the display.	Press — You w	rill see t.F,L on the display.	
21.	Press	until you see	OFF on the display.	Press — You w	rill see In. Sh on the display.	
22.	Press	until you see	OFF on the display.	Press — You w	rt. Lo on the display.	
23.	Press	until you see	10.0 on the display.	Press — You w	rt. H, on the display.	
24.	Press	until you see	120.0 on the display.	Press — You w	vill see tune on the display.	
25.	Press	tab until you return	n to the temp. display.			
26.	Press	until you see	100.0 on the display.	The menu settings are complete.		

ANKOM Technology • 2052 O'Neil Road • Phone: 315-986-8090 • Fax: 315-986-8091