




## Lid O-Ring Replacement A200/A200I



1. Remove the lid from the instrument and soak the lid in acetone in a covered container for three hours.
2. Pry the old O-ring out with a sharp tool.

3. Clean the groove of all old glue residue.
4. Apply the supplied glue on the bottom of the groove.



5. Push the new O-Ring into the groove as shown.
6. Finish pushing the O-ring into the groove until the O-ring is the same height all around the groove.
7. Re-attach the lid to the instrument.